

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

OCT 27 2011

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130526</u>
Name <u>Joseph V. Shoemaker</u>		ID# <u>4188</u> Date <u>10-24-11</u>
A	Agency <u>SDA 1 (SAND POINT P.D.)</u>	Phone # <u>907-383-3700</u>
Instrument Location <u>SAND POINT, AK 99661</u>		
Alco S/N <u>3B</u> Target Value <u>.080</u> High Pressure <u>1000</u>		
Alco Test Value Average <u>.075</u> <u>.074</u>		
1 st Alco 2 nd Alco		
Signature <u>Joseph V. Shoemaker</u>		
COB 10/28/11		
(OVER)		

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9 day of Nov, 2011.

Nikki Roth
Nikki Roth, Notary Public
My Commission Expires With Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130526

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

OCTOBER 24, 2011

OPERATOR'S NAME: *COB 10/23/11*
SHOENAKER J V
OPERATOR'S NUMBER: 4188
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L

O.L. #: L
DEPT/AGENCY: SDF1
CASE/REPORT: L
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: ~~130526~~ *38*
COB 10/29/11

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.16 in
ALCO TARGET .077 17:45
BLANK TEST .000 17:46
INTERNAL STANDARD VERIFIED 17:46
ALCO TV 29.16 in .075 17:46
BLANK TEST .000 17:47
SUBJECT SAMPLE *COB 10/29/11* .000 17:48
BLANK TEST .000 17:49
ALCO TV 29.16 in .074 17:49
BLANK TEST .000 17:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

OCTOBER 24, 2011
TIME 17:53

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 42c
BAROMETER: 29.16 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~